



Accreditation Scope

STRATA MANUFACTURING PJSC, NAL 217
Calibration Laboratory, (ISO/IEC 17025:2017)

ALAIN AIRPORT AREA , ,

Issue Date: 08/07/2024

Issue No: 1

Expiry Date: 07/07/2027

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measurement Range	Expanded Measurement Uncertainty (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P)/Client- site
Linear and angular measurements	Caliper	(0.5 to 300) mm	22 µm	SQP-OP-76- 0001_APP-D	P
Linear and angular measurements	Caliper	(>300 to 1000) mm	50 µm	SQP-OP-76- 0001_APP-D	P
Linear and angular measurements	Micrometer	(0.5 to 25) mm	3 µm	SQP-OP-76- 0001_APP-E	P
Linear and angular measurements	Micrometer	(> 25 to 100) mm	6 µm	SQP-OP-76- 0001_APP-E	P
Linear and angular measurements	Dial Gauges	(0.1 to 25) mm	8 µm	SQP-OP-76- 0001_APP-X	P
Linear and angular measurements	Feeler Gauges	(0.05 to 1) mm	6 µm	SQP-OP-76- 0001_APP-C	P
Linear and angular measurements	Depth Micrometer / Gauge	(0.5 to 100) mm	7 µm	SQP-OP-76- 0001_App E	P
Linear and angular measurements	Thickness Gauge / Snap gauge	(0.5 to 100) mm	8 µm	SQP-OP-76- 0001_App C	P
Linear and angular measurements	Metal Ruler	(0.5 to 1000) mm	1 mm	SQP-OP-76- 0001_App F	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

STRATA MANUFACTURING PJSC, NAL 217
Calibration Laboratory, (ISO/IEC 17025:2017)

ALAIN AIRPORT AREA , ,

Issue Date: 08/07/2024

Expiry Date: 07/07/2027

Issue No: 1

Linear and angular measurements	Measuring Tape	(0.5 to 1000) mm	1 mm	SQP-OP-76-0001_AppF	P
Linear and angular measurements	Go No-Go Gauge	(1 to 20) mm	3 µm	SQP-OP-76-0001_AppG	P
Linear and angular measurements	Bevel Protractor	(5 to 360)°	0.06°	SQP-OP-76-0001_App AB	P
Linear and angular measurements	Engineers Square (Linear Measurement)	(0.5 to 300) mm	9 µm	SQP-OP-76-0001_App N	P
Linear and angular measurements	Engineers Square (Angle Measurement)	90°	0.06°	SQP-OP-76-0001_App N	P
Linear and angular measurements	Angle Block, Angle Gauges	5° to 180°	0.06°	SQP-OP-76-0001_App AB	P
Linear and angular measurements	Straight Edge	(1 to 1200) mm	10 µm	SQP-OP-76-0001_App N	P
Linear and angular measurements	3point Micrometer /Bore Gauge	(0.5 to 60) mm	9 µm	SQP-OP-76-0001_App AF	P
Linear and angular measurements	Height Gauge	(1 to 1000) mm	9 µm	SQP-OP-76-0001_App D	P
Linear and angular measurements	Granite Surface Table	(1000 X 2000) mm	15 µm	SQP-OP-76-0001_App AZ	P/S
Linear and angular measurements	Lever Dial Gauge	(0.01 to 1) mm	5 µm	SQP-OP-76-0001_App X	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

STRATA MANUFACTURING PJSC, NAL 217
Calibration Laboratory, (ISO/IEC 17025:2017)

ALAIN AIRPORT AREA , ,

Issue Date: 08/07/2024

Expiry Date: 07/07/2027

Issue No: 1

Thermodynamics and temperature measurements	Oven	(50 to 300) °C	1.3 °C	SQI-OP-76-0001 _App-BC	P
Thermodynamics and temperature measurements	Water Bath	(0 to 100) °C	1.3 °C	SQI-OP-76-0001 _App-BC	P
Thermodynamics and temperature measurements	Incubator/Freezers	(-25 to 80) °C	1.3 °C	SQI-OP-76-0001 _App-BC	P
Thermodynamics and temperature measurements	Furnace	(100 to 300) °C	2 °C	SQI-OP-76-0001 _App-BC	P
Thermodynamics and temperature measurements	Thermocouple	(-30 to 600) °C	0.54 °C	SQP-OP-76-0001 _App R	P
Thermodynamics and temperature measurements	Dial Thermometer	(-30 to 600) °C	0.54 °C	SQP-OP-76-0001 _App R	P
Thermodynamics and temperature measurements	IR Thermometer	(50 to 100) °C	2 °C	SQP-OP-76-0001 _App AD	P
Thermodynamics and temperature measurements	IR Thermometer	(100 to 500) °C	3.9 °C	SQP-OP-76-0001 _App AD	P
Measurement of force, weight and pressure	non-Automatic Weighing Instruments	100 mg to 220 g	12 mg	SQP-OP-76-0001 _App AA	P/S

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

STRATA MANUFACTURING PJSC, NAL 217
Calibration Laboratory, (ISO/IEC 17025:2017)

ALAIN AIRPORT AREA , ,

Issue Date: 08/07/2024

Expiry Date: 07/07/2027

Issue No: 1

Measurement of force, weight and pressure	non-Automatic Weighing Instruments	> (220 to 3000) g	0.05 g	SQP-OP-76-0001_App AA	P/S
Measurement of force, weight and pressure	non-Automatic Weighing Instruments	(> 3 kg to 40) kg	4 g	SQP-OP-76-0001_App AA	P/S
Measurement of force, weight and pressure	Torque Wrenches / Screwdrivers	(0.2 to 25) N.m	1.5 %	SQP-OP-76-0001_App BB	P
Measurement of force, weight and pressure	Torque Wrenches / Screwdrivers	>(25 to 180) N.m	1.5 %	SQP-OP-76-0001_App BB	P
Measurement of force, weight and pressure	Torque Wrenches / Screwdrivers	(>180 to 1100) N.m	1.5 %	SQP-OP-76-0001_App BB	P
Measurement of force, weight and pressure	Pressure & Vacuum Gauges	(-0.95 to 3) bar	19 mbar	SQP-OP-76-0001_App K	P/S
Measurement of force, weight and pressure	Pressure & Vacuum Gauges	(>3 to 350) bar	0.14 bar	SQP-OP-76-0001_App K	P/S
Measurement of force, weight and pressure	Sound Level Meter	(94 & 114) dB	1.2 dB	SQP-OP-76-0001_App BA	P
Measurement of force, weight and pressure	Stopwatch/Timer	10 sec to 2 hr	2.2 Sec	SQP-OP-76-0001_App V	P
END					

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.