



Accreditation Scope

Voyage Automation Oil Well Equipment Installation LLC, NAL 177
Calibration Laboratory, (ISO/IEC 17025:2017)

M44, Mussafah Industrial Area, Abudhabi - UAE

Issue Date: 01/09/2022

Expiry Date: 31/08/2025

Issue No: 01

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measuring Range	CMC (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P) / Client-site (S)
Temperature	Temperature Sensors/Thermocouples	-20 to 100°C	0.51 °C	EURAMET cg-11 (Ver 2.0): 2011	P
Pressure	Pneumatic Pressure Gauge	-1 to 20 Bar	0.03 bar	DKD R-6-1:2014	P
	Hydraulic Pressure Gauge	0 to 700 Bar	0.24 bar		
END					