



Accreditation Scope

Intertek Fujairah FZC, NAL 227
Testing Laboratory, (ISO/IEC 17025:2017)

Ware House No. 201 & 203 Fujairah Free Zone - II Fujairah, Fujairah, United Arab Emirates

Issue Date: 23/12/2024

Expiry Date: 22/12/2027

Issue No: 1

Testing Field	Materials/ Products Tested	Type of test/ Test parameter/ Properties measured/Range of measurement	Test Method (Standard, Internal Procedure, Technique)	Permanent lab (P) / Client-site (S)
Chemical	Residual Fuel Oil	Kinematic Viscosity at 50°C	ASTM D445-24, IP71-2020, ISO 3104:2020	Lab1
Chemical	Residual Fuel Oil	Density at 15°C	IP 365-2020, ISO 12185:2024, ASTM D4052-22	Lab1
Chemical	Residual Fuel Oil	Water Content	ASTM D95-23, IP 74-2020, ISO 3733:1999	Lab1
Chemical	Residual Fuel Oil	Ash Content at 775°C	IP 4-2012, ASTM D482-19, ISO 6245:2001	Lab1
Chemical	Residual Fuel Oil	Micro Carbon Residue	ASTM D4530- 15(2020)	Lab1
Chemical	Residual Fuel Oil	Sulphur Content	ASTM D4294-21	Lab1
Chemical	Residual Fuel Oil	Total Sediment Accelerated	IP 390-B-11 (2017), ISO 10307-2:2009, ASTM D4870-22	Lab1

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Intertek Fujairah FZC, NAL 227
Testing Laboratory, (ISO/IEC 17025:2017)

Ware House No. 201 & 203 Fujairah Free Zone - II Fujairah, Fujairah, United Arab Emirates

Issue Date: 23/12/2024

Expiry Date: 22/12/2027

Issue No: 1

Chemical	Residual Fuel Oil	Total Sediment Potential	IP 390-A-11(2017), ISO 10307-1:2009. ASTM D4870-22	Lab1
Chemical	Residual Fuel Oil	Pour Point	IP 15-2022, ASTM D97-17b (2022), ISO 3016:2019	Lab1
Chemical	Residual Fuel Oil	Flash Point	IP 34-2021, ASTM D93-20, ISO 2719:2016	Lab1
Chemical	Residual Fuel Oil	Aluminum	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Silicon	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Vanadium	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Sodium	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Iron	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Phosphorus	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Calcium	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Nickel	IP 501-05 (2019)	Lab1

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Intertek Fujairah FZC, NAL 227
Testing Laboratory, (ISO/IEC 17025:2017)

Ware House No. 201 & 203 Fujairah Free Zone - II Fujairah, Fujairah, United Arab Emirates

Issue Date: 23/12/2024

Expiry Date: 22/12/2027

Issue No: 1

Chemical	Residual Fuel Oil	Zinc	IP 501-05 (2019)	Lab1
Chemical	Residual Fuel Oil	Acid Number	ASTM D664-24	Lab1
Chemical	Residual Fuel Oil	Hydrogen Sulphide	IP 570-21	Lab1
Chemical	Residual Fuel Oil	CCAI	ISO 8217:2024 Annex F	Lab1
Chemical	Gas Oil (DMA)	Colour / Appearance	ASTM D4176-22	Lab1
Chemical	Gas Oil (DMA)	Kinematic Viscosity at 40 °C	ASTM D445-24, IP 71-2020, ISO 3104:2020	Lab1
Chemical	Gas Oil (DMA)	Density at 15 °C	IP365-202020, ISO 12185:2024, ASTM D4052-22	Lab1
Chemical	Gas Oil	Density@20 Degree C	ASTM D4052-22	Lab1
Chemical	Gas Oil (DMA)	Water Content	ASTM D95-13(2018), IP 74-2014, ISO 3733:1999	Lab1
Chemical	Gas Oil (DMA)	Ash Content at 775°C	IP 4-2012, ASTM D482-19, ISO 6245:2001	Lab1
Chemical	Gas Oil (DMA)	Micro Carbon Residue	ASTM D4530-15 (2020)	Lab1

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Intertek Fujairah FZC, NAL 227
Testing Laboratory, (ISO/IEC 17025:2017)

Ware House No. 201 & 203 Fujairah Free Zone - II Fujairah, Fujairah, United Arab Emirates

Issue Date: 23/12/2024

Expiry Date: 22/12/2027

Issue No: 1

Chemical	Gas Oil (DMA)	Sulphur Content	ASTM D4294-21, ASTM D 5453-19a	Lab1
Chemical	Gas Oil (DMA)	Cold Filter Plugging Point	IP 309-2016, ASTM D6371-24	Lab1
Chemical	Gas Oil (DMA)	Pour Point	IP 15-2022, ASTM D97-17b (2022), ISO 3016:2019	Lab1
Chemical	Gas Oil (DMA)	Flash Point	IP 34-2021, ASTM D93-20, ISO2719:2016	Lab1
Chemical	Gas Oil (DMA)	Acid Number	ASTM D664-24	Lab1
Chemical	Gas Oil (DMA)	Hydrogen Sulphide	IP 570-2021	Lab1
Chemical	Gas Oil (DMA)	Cetane Index	ASTM D4737-21	Lab1
Chemical	Gas Oil (DMA)	Distillation	ASTM D86-23ae1	Lab1
Chemical	MOGAS	Density at 20°C	ASTM D4052-22	Lab1
Chemical	MOGAS	Sulphur Content	ASTM D5453-19a	Lab1
Chemical	MOGAS	Distillation	ASTM D86-23ae1	Lab1
Chemical	MOGAS	Octane Number, Research	ASTM D2699-24a	Lab1
Chemical	MOGAS	Octane Number, Motor	ASTM D2700-24b	Lab1

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Intertek Fujairah FZC, NAL 227
Testing Laboratory, (ISO/IEC 17025:2017)

Ware House No. 201 & 203 Fujairah Free Zone - II Fujairah, Fujairah, United Arab Emirates

Issue Date: 23/12/2024

Expiry Date: 22/12/2027

Issue No: 1

Chemical	MOGAS	Vapor Pressure of Petroleum Products and Liquid Fuels (Mini Method)	ASTM D5191-22	Lab1
Chemical	MOGAS	Copper Corrosion	ASTM D130-19	Lab1
Chemical	MOGAS	Oxidation Stability	ASTM D525-12a(2019)	Lab1
Chemical	Crude Oil	Density at 15°C	ASTM D5002-22	Lab1
Chemical	Crude Oil	Kinematic Viscosity at 40 °C	ASTM D445-24, IP 71-2020, ISO 3104:2020	Lab1
Chemical	Crude Oil	Sulphur Content	ASTM D4294-21	Lab1
Chemical	Crude Oil	Sediment by Extraction	ASTM D473-22	Lab1
Chemical	Crude Oil	B.S&W	ASTM D4007-22	Lab1
Chemical	Crude Oil	Water Content	ASTM D4006-22	Lab1
END				

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.