



Accreditation Scope

GULF COAST INSTRUMENTATION AND CALIBRATION EQUIPME, NAL 241
Calibration Laboratory, (ISO/IEC 17025:2017)

**OFFICE NO 8, MW5-MUSAFFAH INDUSTRIAL AREA, OPPOSITE TO FIBREX CONSTRUCTION-
ABUDHABI-UAE, Abu Dhabi, United Arab Emirates**

Issue Date: 04/09/2025

Expiry Date: 03/09/2028

Issue No: 1

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measurement Range	Expanded Measurement Uncertainty (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P)/Client- site
Pressure	Pressure Gauge (Digital/Analogue)	1 to 40 bar	0.05 bar	DKD-R 6-1 (03/2014)	P
Pressure	Pressure Gauge (Digital/Analogue)	(>= 40 to 900) bar	0.4 bar	DKD-R 6-1 (03/2014)	P
END					

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.