



Accreditation Scope

GULF AIRCRAFT & ENGINEERING SERVICES (GAES) (FZE), NAL 257
Calibration Laboratory, (ISO/IEC 17025:2017)

600 M2 Warehouse P6-025 P.O.Box 513783 Sharjah - U.A.E, Sharjah, United Arab Emirates

Issue Date: 24/09/2026

Expiry Date: 23/09/2029

Issue No: 1

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measurement Range	Expanded Measurement Uncertainty (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P)/Client- site
Measurement of force, weight and pressure	Pressure / Vacuum Pressure Meters, Pressure Sensors, Pressure Transducers	-0.9 bar to 10 bar	0.005 bar	CAL-PRE- 01:2026	P/S
Measurement of force, weight and pressure	Pressure / Vacuum Pressure Meters, Pressure Sensors, Pressure Transducers	0 to 40 bar	0.017 bar	CAL-PRE- 01:2026	P/S
Measurement of force, weight and pressure	Pressure / Vacuum Pressure Meters, Pressure Sensors, Pressure Transducers	0 to 400 bar	0.3 bar	CAL-PRE- 01:2026	P/S
Measurement of force, weight and pressure	Pressure / Vacuum Pressure Meters, Pressure Sensors, Pressure Transducers	>400 bar to 1000 bar	0.4 bar	CAL-PRE- 01:2026	P/S

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

GULF AIRCRAFT & ENGINEERING SERVICES (GAES) (FZE), NAL 257
Calibration Laboratory, (ISO/IEC 17025:2017)

600 M2 Warehouse P6-025 P.O.Box 513783 Sharjah - U.A.E, Sharjah, United Arab Emirates

Issue Date: 24/09/2026

Expiry Date: 23/09/2029

Issue No: 1

Measurement of force, weight and pressure	Torque Wrench, Torque Transducer, Torque screwdriver, Torque Calibrator	0 to 5 Nm	1% of rdg	CAL-TW-01:2026	P
Measurement of force, weight and pressure	Torque Wrench, Torque Transducer, Torque screwdriver, Torque Calibrator	0 to 50 Nm	1% of rdg	CAL-TW-01:2026	P
Measurement of force, weight and pressure	Torque Wrench, Torque Transducer, Torque screwdriver, Torque Calibrator	0 to 500 Nm	0.5% of rdg	CAL-TW-01:2026	P
Measurement of force, weight and pressure	Torque Wrench, Torque Transducer, Torque screwdriver, Torque Calibrator	0 to 2000 Nm	0.4% of rdg	CAL-TW-01:2026	P
Measurement of force, weight and pressure	Dynamometers/Force Gauges / Aircraft Weighing Scale	0 to 30 ton	0.04% of rdg	CAL-FOR-01:2026	P
Measurement of force, weight and pressure	Dynamometers/Force Gauges / Load Cell / (Compression)	0 to 250 kN	0.2% of rdg	CAL-FOR-02:2026	P
Linear and angular measurements	Protractor	0° to 180°	0.5°	CAL-DIM-12:2026	P
Linear and angular measurements	Caliper	0.5 mm to 200 mm	0.01 mm	CAL-DIM-01:2026	P/S

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

GULF AIRCRAFT & ENGINEERING SERVICES (GAES) (FZE), NAL 257
Calibration Laboratory, (ISO/IEC 17025:2017)

600 M2 Warehouse P6-025 P.O.Box 513783 Sharjah - U.A.E, Sharjah, United Arab Emirates

Issue Date: 24/09/2026

Expiry Date: 23/09/2029

Issue No: 1

Linear and angular measurements	Caliper	> 200 mm to 600 mm	0.02 mm	CAL-DIM-01:2026	P/S
Linear and angular measurements	Inside micrometer	0 to 100 mm	0.005 mm	CAL-DIM-03:2026	P
Linear and angular measurements	Dial Gauge	0 to 50 mm	0.006 mm	CAL-DIM-11:2026	P/S
Linear and angular measurements	Optical Micrometer	0 to 36 mm	0.01 mm	CAL-DIM-15:2026	P/S
Linear and angular measurements	Depth gauge	0.5 mm to 200 mm	0.01 mm	CAL-DIM-09:2026	P/S
Linear and angular measurements	Depth gauge	> 200 mm to 300 mm	0.02 mm	CAL-DIM-09:2026	P/S
Linear and angular measurements	Thread Gauge	50 mm (0 to 90) °	0.004 mm 0.009°	CAL-DIM-17:2026	P
Linear and angular measurements	Telescopic gauge Set	0 to 100 mm	0.003 mm	CAL-DIM-16:2026	P
Linear and angular measurements	Bore Gauge	0 to 150 mm	0.006 mm	CAL-DIM-10:2026	P
Linear and angular measurements	Feeler gauge	0 to 25 mm	0.0028 mm	CAL-DIM-08:2026	P/S
Linear and angular measurements	Go No Go gauge	0 to 25 mm	0.0025 mm	CAL-DIM-07:2026	P/S
Linear and angular measurements	Straight edge	1000 mm	0.005 mm	CAL-DIM-05:2026	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

GULF AIRCRAFT & ENGINEERING SERVICES (GAES) (FZE), NAL 257
Calibration Laboratory, (ISO/IEC 17025:2017)

600 M2 Warehouse P6-025 P.O.Box 513783 Sharjah - U.A.E, Sharjah, United Arab Emirates

Issue Date: 24/09/2026

Expiry Date: 23/09/2029

Issue No: 1

Linear and angular measurements	Measuring tape	5 m	0.01 mm	CAL-DIM-06:2026	P
Linear and angular measurements	Measuring scale	0 to 1000 mm	0.01 mm	CAL-DIM-06:2026	P
Linear and angular measurements	Outside micrometer	0 to 300 mm	0.002 mm	CAL-DIM-02:2026	P/S
Linear and angular measurements	Ultrasonic thickness gauge	0 to 250 mm	0.0088 mm	CAL-DIM-13:2026	P/S
Linear and angular measurements	Radius gauge	100 mm	0.0046 mm	CAL-DIM-04:2026	P
Linear and angular measurements	Grip gauge	50 mm	0.01 mm	CAL-DIM-14:2026	P
END					

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.